



9410 - 20 Ave N.W.
Edmonton, Alberta, Canada T6N 0A4
Tel: (780) 437-9100 / Fax: (780) 437-7787

September 18, 2006

Attention: Jason W. Kuzmiak
OHMART/VEGA CORPORATION
4241 ALLENDORF DRIVE
CINCINNATI, OH 45209

Re: Design registration

The design submission, tracking number 2006-06804, originally received on August 09, 2006 was surveyed and accepted for registration as follows:

CRN : 0F04145.2 **Accepted on:** September 18, 2006
Reg Type: Addition to Acc. Fitting **Expiry Date:** September 10, 2015
Drawing No. : GE2255, GE2256 & D-V-2230 As Noted
Fitting type: LEVEL SENSORS - VEGAVIB 60 & VEGAWAVE 60 SERIES

The registration is conditional on your compliance with the following notes:

This registration has been issued based on the understanding that:

- * It includes the VegaVIB 61, 62 & 63 Series and Vega WAVE 61, 62 & 63 Series;
- * The connection flange is made of A/SA-182 F316L;
- * The maximum design pressure for the sensor is 1600 kPa.

An acceptance letter and a copy of the stamped drawing with Statutory Declaration will be sent to you.

An invoice covering survey and registration fees will be forwarded from our Revenue Accounts.

If you have any question don't hesitate to contact me by phone at (780) 433-0281 ext 3330 or fax (780) 437-7787 or e-mail Preda@absa.ca.

Sincerely,

A handwritten signature in cursive script that reads 'Preda'.

PREDA, DRAGOS, P. Eng.
Design Survey Engineer