



Ref. Certif. No.

CA/22158/CSA

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

VEGATRENN Series Single- and double channel separators

Name and address of the applicant

VEGA Grieshaber KG
Am Hohenstein 113, 77761 Schiltach, GERMANY

Name and address of the manufacturer

VEGA Grieshaber KG
Am Hohenstein 113, 77761 Schiltach, GERMANY

Name and address of the factory

VEGA Grieshaber KG
Am Hohenstein 113, 77761 Schiltach, GERMANY

Note: When more than one factory, please report on page 2

Additional Information on page 2

Ratings and principal characteristics

Refer to page 2 of Certificate.

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

N/A

Model / Type Ref.

VEGATRENN 141, 142, 151, 152

Additional information (if necessary may also be reported on page 2)

This certificate replaces certificate CA/19977/CSA dated 2017-03-21.

Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010, IEC 61010-2-201:2013, refer to CB Report for National Deviations.

As shown in the Test Report Ref. No. which forms part of this Certificate

CB 153857-70076786 (70199783)

This CB Test Certificate is issued by the National Certification Body



CSA Group

CSA Group
178 Rexdale Boulevard
Toronto, ON M9W 1R3 Canada

Date: 2018-11-07

Signature: Vegard Andersen





Ref. Certif. No.

CA/22158/CSA

Ratings and principal characteristics:

VEGATRENN 141, 142:

24...230 Vac, 50/60 Hz, < 5 W, < 20 VA

24...65 Vdc, < 5 W, < 20 VA

VEGATRENN 151, 152:

Passive, no extra voltage supply: 15...35 Vdc, 4...20 mA

Additional information (if necessary)

Date: 2018-11-07

Signature: Vegard Andersen

