

VEGADIF 85 Profibus PA

Version, available since	Description	Device Rev.
1.2.2, 06/2020	Error correction: <ul style="list-style-type: none"> – PA communication <ul style="list-style-type: none"> – Access rights adapted to requirements in profile 3.02: <ul style="list-style-type: none"> – Parameter SCALE_OUT und PRIMARY_VALUE_TYPE must be writable – I&M data must be readable only 	3
1.2.1, 11/2018	Error correction: <ul style="list-style-type: none"> – Measurement function: <ul style="list-style-type: none"> – After triggering an automatic position correction and a restart of the sensor, it could happen that the device went into error state F261 – PLICSCOM adjustment: <ul style="list-style-type: none"> – Depending on the units set, the limit values in the position correction menus were displayed incorrectly 	3
1.2.0, 06/2018	First version New functions and modifications relating to VEGADIF 65: <ul style="list-style-type: none"> – Measurement function: <ul style="list-style-type: none"> – Increased accuracy – Quicker reaction time – Extension with application parameter adjustment – Instrument software, in general: <ul style="list-style-type: none"> – Lower supply voltages possible – Device status according to NE 107 – Event memory added – Function extension for the measured value memory – Real time clock added – PLICSCOM adjustment: <ul style="list-style-type: none"> – Modification of the menu structure – Modification of the layout with value changes – The following languages are available: <ul style="list-style-type: none"> – German – English – French – Spanish – Russian – Italian – Dutch – Portuguese – Czech – Polish – Turkish – Chinese – Japanese 	3

Service info plics® software versions



Version, available since	Description	Device Rev.
	<ul style="list-style-type: none"> – PA communication – Profibus PA Profile 3.02 – 3 AI function blocks available 	

Legend:

Name	Description
Version	Compatibility version.Function extension version.Error correction version
available since	Month/Year
Device Rev.	Version number of the instrument defined by HART. Consecutive integral number Will be increased if in the "Application Layer" modifications were carried out, e.g. new commands, modifications in the data structure in a command.